

Specifications

Parameters	Detail
Detection Method	Real time PCR
Detection Capability	2 PCR Chambers 4 Fluorescent Probe Channels (FAM, JOE, ROX, CY5)
Instrument Module Coupling	Up to 8 modules with Console
Average Heating Rate	From 50°C to 90°C, ≥13°C/s
Average Cooling Rate	From 90°C to 50°C, ≥10°C/s
Temperature Accuracy	≤±0.5°C
Data Interfaces	USB*2、CAN、COM、LAN、WiFi
Data Management	Build -in printer Bidirectional LIS connection
Power Requirements	AC100V-240V, 50/60Hz
Weight	Console: 4.5kg Module: 14kg
Dimensions	Console: 200mm×388mm×214mm(L×W×H) Module: 345mm×150mm×403mm(L×W×H)
Catalogue Number	WK46/WK47/WK48/WK49/WK50/WK51/WK52/WK53

Racing For Life

Guangzhou Wondfo Biotech Co., Ltd.
No.8 Lizhishan Road, Science City,
Huangpu District, 510663, Guangzhou, P.R.China
Tel: (+86) 400-830-8768
Website: en.wondfo.com
E-mail: sales@wondfo.com.cn



© 2025 Guangzhou Wondfo Biotech Co., Ltd.
All rights reserved Rev.A3 Rel.: 2026/05/16

Wondfo

U-CardDx 200

Automated Rapid Molecular Diagnostic System

Redefining PCR with simplicity, Efficiency, Precision

15 minutes
result output

7 targets in
one cartridge

8 modules
maximumly



Guangzhou Wondfo Biotech Co., Ltd.

Features



One-step operation

- Sample in , result out



Powerful report capability

- 7 targets can be detected in one cartridge
- Extendable test panel
- Result output in 15 minutes



Scalable modular configuration

- Multi channel configuration :
- 8 units maximumly, 256T per day



Allowing various specimens

- Alveolar lavage fluid, Sputum, Nasopharyngeal swab, Oropharyngeal swab, Cervical swab, Urine, Blood, Feces



All in one mini lab design simplifying working procedures

- Storage at room temperature
- Dry and liquid reagents prefilled



Test Procedure



1 Scan the sample ID



2 Scan the cartridge code



3 Load sample



4 Insert the cartridge



5 Various Displays of Results

Test Menu

Product Name	Detection Target	Specimen Type	Catalogue Number
2019-nCoV/FluA/FluB/RSV Test	2019-nCoV,Flu A,Flu B,RSV	Nasopharyngeal swab, Oropharyngeal swab	WR38P0004 (8T), WR38P0005 (24T), WR38P0006 (48T)
Respiratory Pathogen Panel	2019-nCoV,Flu A,Flu B, RSV,MP,Adv,HRV	Nasopharyngeal swab, Oropharyngeal swab	WR39P0004 (8T), WR39P0005 (24T), WR39P0006 (48T)
Respiratory Pathogen Panel Test(B)	2019-nCoV, Flu A, Flu B, RSV, Strep A, Adv, HRV	Nasopharyngeal swab, Oropharyngeal swab	WR94P0001 (8T), WR94P0002 (24T), WR94P0003 (48T)
Strep A Test	speB gene	Oropharyngeal swab	WR91P0001 (8T), WR91P0002 (24T), WR91P0003 (48T)
Lower Respiratory Pathogen Panel Test	HI,SP,SA,KPN,BP,PA,LP	Sputum	WR40P0004 (8T), WR40P0005 (24T), WR40P0006 (48T)
Vector-borne Pathogen Panel Test	DENV1-4,CHIKV,ZIKV,WNV	Serum, Plasma	WR41P0010 (8T), WR41P0011(24T), WR41P0012 (48T)
CT/NG Test	CT,NG	Urine, Cervical or vaginal swab	WR43P0004 (8T), WR43P0005 (24T), WR43P0006 (48T)
STI Plus Test	CT,NG,UU,MG,TV,MH,HSV	Urine, Cervical or vaginal swab	WR44P0001 (8T), WR44P0002 (24T), WR44P0003 (48T)
GBS Test	GBS (sip,cfb)	Vaginal and rectal swabs	WR45P0004 (8T), WR45P0005 (24T), WR45P0006 (48T)
HPV Test	16,18, 45,11HR HPV	Cervical cells	WR48P0004 (8T), WR48P0005 (24T), WR48P0006 (48T)
MTB Test	IS6110/IS1081 gene	Sputum, Sputum sediment	WE44P0001 (8T), WE44P0002 (24T), WE44P0003 (48T)
MTB&RIF Test	MTB (IS6110/IS1081), RIF (rpoB)	Sputum	WR46P0007 (8T), WR46P0008 (24T), WR46P0009 (48T)
MTB INH&FQs Test	INH (katG/inhA/OxyR-ahpC), FQs (gyrA/gyrB)	Sputum, Culture isolates	WR58P0004 (8T), WR58P0005 (24T), WR58P0006 (48T)
MRSA/SA Test	SA (SPA/nuc), MRSA (mecC/mecA, SCCmec)	Nasal swabs, skin or soft tissue infection swabs	WR42P0004 (8T), WR42P0005 (24T), WR42P0006 (48T)
Carba Test	blaKPC, blaNDM, blaVIM, blaOXA-48, blaOXA-23, blaIMP	Rectal or Perirectal Swab, Bacterial Isolates	WR88P0001 (8T), WR88P0002 (24T), WR88P0003 (48T)
VRE Test	VRE (vanA,vanB)	Rectal swab, Anal swab	WR47P0004 (8T), WR47P0005 (24T), WR47P0006 (48T)
C. difficile Test	toxin B gene	Unformed stool	WR89P0001 (8T), WR89P0002 (24T), WR89P0003 (48T)
GI-Virus Panel	NoV GI, NoV GII, RVA, EAdV, SaV, AstV	Unformed stool	WR90P0001 (8T), WR90P0002 (24T), WR90P0003 (48T)

U-CardDx 200 Cartridge

- 1 minute hands-on preparation
- Delivery of qualitative and quantitative results
- More accurate and sensitive: LoD 200 copies/mL
- Lyophilized reagents for stability

